Endossage® TP



Contents

Product Assistance	3
Warnings and Disclaimers	4
Purpose and Indications	5
Product Description	6
Usage Environment	7
Endossage TP Set-Up	8
How to Operate Endossage TP	14
Endossage TP Product Care	19
Troubleshooting	25
Product Quick Guide	26

Product Assistance

The makers of Endossage TP are committed to the safe effective use of our products.

If you need any assistance in the operation of or have questions about our product, please contact your distributor or contact us directly at:

Info@gremedix.com

www.endossage.com

888-748-7077

Please have your product information ready including date and place of purchase along with your product serial number so we can more easily assist you with specific product concerns or questions.



Warnings and Disclaimers



Endossage TP is not currently an FDA approved or classified medical device. The makers of Endossage make no specific claims and can not guarantee the efficacy or therapeutic value of Endossage.

Endossage is in the process of FDA device classification and therapeutic trial studies in multiple locations. Check our website for new information regarding updates to the device studies and FDA classification and approval status.

Endossage is not currently a prescription device, however the manufacturer strongly recommends users of the device consult their physician, therapist, or qualified medical professional before using Endossage. The manufacturer holds no liability regarding the use of Endossage for any purpose.

Endossage should be used only as directed in this manual and in conjunction with the advice and training of a qualified medical professional.



If at any time while using Endossage TP you experience pain, discomfort, or ANY adverse reaction, STOP using the product immediately! Contact your medical professional for advice. If you suspect a product defect or functional problem contact your distributor or contact us directly at:

Info@gremedix.com or 888-748-7077

Please have your product information ready including date and place of purchase along with your product serial number so we can more easily assist you with specific product concerns or questions.

Purpose and Indications

Endossage TP was developed to augment physical therapy treatments in the area of pelvic pain.

Endossage TP is designed to compliment such treatments that may involve the use of dilation, manipulation or massage, or trigger point therapies for a wide range of indictions.

Individuals that have used Endossage TP have done so for the following list of indications:

- Painful Sexual Intercourse
- General Vaginal Pain
- High Tone Pelvic Floor Dysfunction
- Vaginismus
- Dyspareunia
- Vestibulodynia
- Vaginal Atrophy
- Vaginal Agenesis
- Vulvar Dermatoses
- Port Radiation Adhesions
- Pudendal Neuralgia

Potential indications may also include conditions stemming from postpartum vaginal trauma, post surgical discomfort from prolapse or bladder mesh surgeries, as well as other uro-gynecological situations.

Check with your qualified medical professional, physician, or physical therapist to discuss if Endossage TP might be right for you.

Product Description

Endossage TP consists of two components.

Endossage TP Base Control Unit includes:

Handle with Control Button and Battery Housing Flexible Neck Attachment Connector Safety and Storage Cap



Interchangeable Trigger Point Attachment



Usage Environment

Endossage TP is an electromechanical device designed for use in a clinical or home use environment.

Endossage should be operated only in a room temperature climate controlled dry environment.

*****CAUTION SHOCK HAZARD*****



Endossage TP should never be operated in extreme and/or outdoor climates or in areas where the device may become wet from the elements or areas which might allow splashing of water or immersion in water such as a bathroom or pool area.



Never immerse Endossage TP in water or other liquid. Do not wash Endossage TP Base Control Units under running water as it NOT water proof and will damage the electronic components.

Endossage TP should be used with the advice and recommendation of a qualified medical professional, physician, or physical therapist.

Endossage TP Set-Up

Setting up Endossage TP is a simple process.

After removing Endossage TP from the package, be sure you have the following items:

- Endossage TP Base Control Unit
- Trigger Point Attachment
- 1. Remove the storage cap from a selected attachment.

To remove the storage cap:

Hold the base of the attachment and the storage cap as shown.



While applying slight pressure on the cap toward the attachment, rotate the cap counter-clockwise. The cap will disengage from the attachment and can be completely removed.



- 2. Inspect the attachment for any damage to the silicone cover or connecting surface.
 - If damaged in any way STOP DO NOT USE.



3. Remove the storage and safety cap from the Base Control Unit

To remove the safety cap:

Hold the Base Control Unit at the attachment connector/ motor drive housing with one hand and the safety cap with the other as shown.



While applying slight pressure on the cap toward the attachment connector housing, rotate the cap counterclockwise. The cap will disengage from the attachment and can be completely removed.



4. Connect the attachment to Base Control Unit.

To connect the attachment:

Hold the base of an attachment in one hand and the Base Control Unit at the attachment connector housing in the other as shown.



Carefully position the attachment such that the small white electrical plug of the attachment lines up with the small white electrical socket of the control unit if so equipped.



Notice that the attachment has three locking tabs that are guided by three corresponding slots on the attachment connector housing as you seat the attachment.



Gently slide the attachment in to the connector housing such that attachment seats down enough for the red O-ring to contact both the attachment connector housing rim and attachment lip.





Never operate the unit without the O-ring in place. Doing so will possibly damage the unit and may cause the attachment to disconnect while in use.

Rotate the attachment clockwise. The attachment will engage the attachment connector housing. You will feel and hear a slight click as it locks into place.



Endossage TP is now ready for use.



We strongly recommend covering the attachment portion with a probe cover or condom of choice to protect the attachment and promote cleanliness and hygiene and to decrease the risk of infection or adverse contact reaction.



How to Operate Endossage TP


How you use Endossage TP is a subject best discussed with your medical professional. Below are instruction on how to use the controls and features of the device to achieve the therapy specified in your particular case.

After initial set-up, operating an Endossage unit is a simple process.

Operating the controls:

Power On/Off Button

To turn on the optional vibration motion of your device press red button until it clicks and stays in the on position. This should start the vibration motor and you should feel a mild vibration. To turn off the optional vibration, press the red button again.



How to Operate Endossage TP - con't

Flexible Neck

Endossage TP has a flexible neck which allows the user to articulate and position the attachment with relation to the base control unit in such a way that the device use is most comfortable and relaxing for the user.

It is important for the user to properly handle and position the flexible neck so as not to damage parts of the device and also achieve the best positioning

To adjust the flexible neck, grasp the device with two hands. Place one hand on the end of the flexible neck closest to the base control unit. Place the other hand at the other end of the flexible neck at the base of the attachment connector assembly as shown.



How to Operate Endossage TP - con't


Holding firmly, manipulate the neck to a shape or position you require. The neck is moderately stiff so as to keep the shape desired and may require some effort to articulate, however excessive force should never be used or required.

If needed you can place your hands on any points along the flexible neck as well to achieve the required shape.



NEVER hold the device by the attachment to bend the flexible neck. Force used to bend the neck while holding by the attachment may break the attachment!







If you have difficulty manipulating the flexible neck, check for damage to the device. If the device appears defective, contact your distributor or contact us for more information. If the product appears sound, speak to your medical professional for more guidance.

Endossage TP Product Care

Endossage TP requires minimal maintenance and care if used as directed. Proper cleaning and storage both extend the life of the product and reduce the risk of infection or reactions from bacteria or other contaminants that may come in contact with the device.

Cleaning

Base Control Unit

Before cleaning, be sure the safety and storage cap along with the included red O-ring are installed on the attachment connector housing of the device. Failure to install the cap may result in liquid contaminating the electronics and mechanism resulting in poor product performance or product failure.

GreMedix recommends using bleach based disinfecting wipes on the exterior components of the base control unit and flexible neck portions of the device for cleaning and disinfecting purposes. Clean the unit with the wipes to remove any visible dirt or contaminants. Allow the unit to dry completely before storage or use.

Do NOT use harsh chemicals such as ammonia, concentrated lye, glass cleaners or household cleaners. These substances contain chemicals that may degrade the surface finish of the product and may affect physical properties of the device. Isopropyl alcohol may be used only on a very limited basis and will degrade the product integrity over time as well.

Attachments

Before cleaning, be sure the storage cap along with the included red O-ring are installed on the connector end of the attachment. Failure to install the cap may result in liquid contaminating the mechanism resulting in poor product performance or product failure.

To remove surface dirt and particles, the attachment can be cleaned with regular soap and water while the storage cap is in place. The attachment should be cleaned and rinsed thoroughly and allowed to air dry.

To disinfect the attachment the manufacturer recommends bleach based disinfecting wipes on the silicone and plastic surfaces. After disinfecting with the wipes, we recommend rinsing the attachment thoroughly with clean running water to minimize any irritation that may be cause by the disinfecting solution. The attachment should be then left to air dry before storage or use.

Do NOT use harsh chemicals such as ammonia, concentrated lye, glass cleaners or house hold cleaners. These substances contain chemicals that may degrade the surface finish of the product and may affect physical properties of the device. Isopropyl alcohol may be used only on a very limited basis and will degrade the product integrity over time as well.

Always inspect the silicone surface of the attachments for nicks or cuts. If any are found, DO NOT USE the attachment. Discard the attachment and use a new one.

Storage

Base Control Unit


When not in use, it is recommended that the base control unit and power adapter be stored in the original foam padded box with the storage and safety cap installed. This will help keep the unit clean for future use and offer protection from accidental damage. One attachment may also be stored in the base unit storage box. Additional attachments should be stored in the purchase box with the storage cap installed and in a clean zipper lock storage bag if possible.

To install the safety and storage cap:

Hold the safety and storage cap in one hand and the Base Control Unit at the attachment connector housing in the other as shown.



Notice that the safety and storage cap has three locking tabs that are guided by three corresponding slots on the attachment connector housing to insert the safety and storage cap.



Gently slide the safety and storage cap into the slots such that safety and storage cap seats down enough for the red O-ring to contact the attachment connector housing.

Rotate the attachment clockwise. The safety and storage cap will engage the attachment connector housing. You will feel and hear a slight click as it locks into place.



Attachments


Cleaned attachments should be stored in the purchase box with the storage cap installed and in a clean zipper lock storage bag if possible.

To install the storage cap:

Hold the base of the attachment and the storage cap as shown.



Notice the three locking tabs on the attachment that correspond to three slots on the storage cap. Align the tabs on the attachment with the slots on the cap and insert the attachment until the red O-ring is in contact with both the attachment lip seat and the safety cap rim.



© 2018 GreMedix LLC

While applying slight pressure on the cap toward the attachment, rotate the cap clockwise.

The cap will engage the attachment and can be completely removed. You will feel and hear a slight click as it locks into place.



Troubleshooting

Hard to connect attachment

- Check for damage to connector tabs. If damaged, replace attachment.
- Check for damage to small electrical plug. If damaged, replace attachment.
- Check for debris on connector plug and surfaces of both the adapter and the base unit. Remove debris if found.
- Verify that base attachment connector housing attachment point is in the correct position for attachment alignment.
- Ensure O-ring is in place

Attachment will not disconnect

- Ensure the attachment is rotated fully counter-clockwise before attempting removal. Rotate fully.
- Wiggle attachment gently while removing. Some slight force is required at times due environmental conditions causing material expansion and contraction.

Attachment Optional Vibration Not Funtioning

- Be sure the attachment is equipped with vibration option.
- Ensure attachment is fully engaged to base unit with O-ring in place. Install O-ring if missing and reattach.
- Be sure power button is fully depressed to on position.
- Check for damage to attachment's white electrical connector.
- Possible dead batteries. Replace with 2 new LR44 button batteries.

Product Quick Guide

Removing Safety and Storage Cap (page 9)

Depress gently and rotate counter-clockwise.

Installing Attachment (page 10)

- 1. Align attachment locking tabs and electrical connector with base connector slots and electrical socket.
- 2. Slide attachment down drive shaft guided by tabs, slots and electrical connector until attachment O-ring is in contact with connector housing rim.
- 3. Gently press attachment toward control unit attachment housing while rotating clockwise until the attachment clicks into place.

If you need any assistance in the operation of or questions about our product, please contact your distributor or contact us directly at:

> Info@gremedix.com www.endossage.com 888-748-7077



Endossage should be used only as directed in this manual and in conjunction with the advice and training of a qualified medical professional.

